



NATIONAL TYPE EVALUATION PROGRAM

Certificate of Conformance

for Weighing and Measuring Devices

For:

Load Cell
Digital, Bending Beam
Model: BL and BL-Ex
 n_{max} : 4000, Single and Multiple Cell, Class III
Capacity: See table below

Submitted By:

EILERSEN ELECTRIC DIGITAL SYSTEMS A/S
Kokkedal Industripark 4
Kokkedal, Denmark DK-2980
Tel: +45 49 180 100
Contact: Frederik Eilersen
Email: info@eilersen.com
Website: www.eilersen.com

Standard Features and Options

Standard Features:

- Digital Communications
- Maximum Excitation – 24VDC
- High overload tolerance (>300% Emax)
- RS485 – 2-wire
- Stainless Steel Construction
- Hermetically sealed (IP68)

Model and Capacity Description:

Model	Capacity	Minimum Load Cell Interval (v_{min})		Minimum Dead Load
		Single Cell	Multiple Cell	
BL	20 kg	5 g	5 g	0
BL	30 kg	7.5 g	7.5 g	0
BL	50 kg	12.5 g	12.5 g	0
BL	80 kg	20.0 g	20.0 g	0
BL	100 kg	25 g	25 g	0
BL	150 kg	37.5 g	37.5 g	0
BL	250 kg	62.5 g	62.5 g	0
BL	300 kg	75.0 g	75.0 g	0
BL	500 kg	125 g	125 g	0
BL-Ex	20 kg	5 g	5 g	0
BL-Ex	30 kg	7.5 g	7.5 g	0
BL-Ex	50 kg	12.5 g	12.5 g	0
BL-Ex	80 kg	20.0 g	20.0 g	0
BL-Ex	100 kg	25 g	25 g	0
BL-Ex	150 kg	37.5 g	37.5 g	0
BL-Ex	250 kg	62.5 g	62.5 g	0
BL-Ex	300 kg	75.0 g	75.0 g	0
BL-Ex	500 kg	125 g	125 g	0

This device was evaluated under the National Type Evaluation Program and was found to comply with the applicable technical requirements of *Handbook 44: Specifications, Tolerances and Other Technical Requirements for Weighing and Measuring Devices*. Evaluation results and device characteristics necessary for inspection and use in commerce are on the following pages. *Editorial changes, not affecting the type or metrological content, corrected this certificate.

Ivan Hankins
Chairman, NCWM, Inc.

Hal Prince
Chair, NTEP Committee
Issued: July 1, 2022

1135 M Street, Suite 110 / Lincoln, Nebraska 68508

The National Conference on Weights and Measures (NCWM) does not approve, recommend, or endorse any proprietary product or material, either as a single item or as a class or group. Results shall not be used in advertising or sales promotion to indicate explicit or implicit endorsement of the product or material by the NCWM.



EILERSEN ELECTRIC DIGITAL SYSTEMS A/S
Load Cell / BL and BL-Ex

Application: The load cells may be used in Class III scales for single cell and multiple cell applications consistent with the model designations, number of scale divisions, and parameters specified in this certificate. Load cells of a given accuracy class may be used in applications with lower accuracy class requirements provided the number of scale divisions, the v_{\min} value, and temperature range are suitable for the application. The manufacturer may market the load cell with fewer divisions (n_{\max}) and with greater v_{\min} values than those listed on the certificate. However, the load cells must be marked with the appropriate n_{\max} and v_{\min} for which the load cell may be used.

Identification: A pressure sensitive identification label located on the cell, states the manufacturer name, model number, the load cells serial number, rated capacity, Accuracy Class, and NTEP CC number. Other pertinent information will be specified on the Calibration Certificate accompanying the cell.

Test Conditions: A Model BL, 20 and 100 kg capacity load cells were tested by FORCE Technology at their facility in Denmark. Testing was conducted in accordance with the OIML-CS arrangement, signed by the NCWM as a utilizing participant for load cell testing. Testing was conducted using deadweights as the reference standard. The load cell was tested over a temperature range of -10 °C to 40 °C with tests run at each temperature. The temperature effect on zero was measured and a time dependence (creep) test was performed. The barometric pressure test to determine sensitivity of the load cell design to changes in barometric pressure was conducted. The data was analyzed for single and multiple load cell applications. OIML R60 selection criteria was used to determine which load cell capacities were tested.

Evaluated By: M. L. Sørensen (FORCE Technology, Denmark)

Type Evaluation Criteria Used: *Handbook 44 Specifications, Tolerances, and Other Technical Requirements for Weighing and Measuring Devices*, 2022 Edition. *NCWM Publication 14: Weighing Devices*, 2022 Edition.

Conclusion: The results of the evaluation and information provided by the manufacturer indicate the device complies with applicable requirements.

Information Reviewed By: D. Flocken (NCWM)

Example(s) of Device:



Photo represents both BL and BL-Ex models